

**0.25% Trypsin-EDTA (1X)**  
Trypsin-EDTA

**Catalog Number:** 1932085  
**Lot Number:** 2147251

---

Storage Temperature: -5 to -20C  
Expiration Date: 2023-11-24

---


For Research Use or Further Manufacturing. Not for diagnostic use or direct administration into humans or animals.

---

TEST	SPECIFICATION	RESULT	UNITS
<sup>1</sup> In-Vitro Bioassay	Acceptable	Acceptable	
<sup>2</sup> Mycoplasma Testing <i>Testing is performed on raw material Trypsin prior to use in manufacturing.</i>	Negative	Negative	
<sup>3</sup> Osmolality	>=270 to <=320	292	mOsm/kg
<sup>4</sup> pH	>=7.2 to <=8.0	7.8	
<sup>5</sup> Porcine Parvovirus Testing <i>Testing is performed on raw material Trypsin prior to use in manufacturing.</i>	Negative	Negative	
<sup>6</sup> Sterility Testing	Negative	Negative	

Read SDS.

Raw stock trypsin has been verified for e-beam irradiation and has tested negative for PPV, mycoplasma and PCV 1/2.



Quality Systems Department

Date: 23-Nov-2021

